

Laser safety class test certificate

No.: CTNT230227011-ZS

Applicant : Guangzhou Labsim Biotech Co.,LTD.
Address : Room 425-433, Building H6, No.39 Ruihe Road,Huangpu District,Guangzhou City,Guangdong Province, China
Manufacturer : Guangzhou Labsim Biotech Co.,LTD.
Address : Room 425-433, Building H6, No.39 Ruihe Road,Huangpu District,Guangzhou City,Guangdong Province, China
Product : Tissue Cassette Laser Printer
Trade Mark : N/A
Model(s) : TB686
Laser Class : Class I (Wavelength $355\text{nm} \pm 1\text{nm}$)
Test Report No. : CTNT230227011
Test Standards : 21CFR 1040.10

The EUT described above has been tested by us with the listed standards Code of Federal Regulations Title 21 Chapter I Subchapter J Part 1040.10 "Performance Standards For Light-Emitting Products-Laser products" and found in compliance with the above mentioned laser safety class (For specific technical indicators, please refer to the test report numbered above).

The test results apply only to the particular sample tested and the applicative tests carried out. This statement is based on a single evaluation of sample above mentioned product. It does not imply an assessment of the whole production process.

The test report is part of the test certificate, and the test certificate is valid when used together with the report.

The test certificate can be verified on the website www.ctnt-cert.com.



Shenzhen Zhongwei Testing Technology Co., Ltd.

Add: Room 1A106, 1/F., No.109, Lijia Road, Henggang, Henggang Street,
Longgang District, Shenzhen, Guangdong, China

Web: <http://www.ctnt-cert.com>

Tel: 86-755-2868 0489

Email: admin@ctnt-cert.com

