

## TEST REPORT

REPORT NO.: CTNT240822003R



Product Name: Combo Electrotherapy Device

Model No. : M100A-WH1

Applicant: Shenzhen Roundwhale Technology Co., Ltd.

Test Procedure: Entrustment Test

Shenzhen Zhongwei Testing Technology Co., Ltd.





**TEST REPORT****EN 60335-1****Safety of household and similar electrical appliances**

Report Number..... : CTNT240822003R

Date of issue..... : Sep.27, 2024

**Name of Testing Laboratory  
preparing the Report .....**

Shenzhen Zhongwei Testing Technology Co., Ltd.  
Room 1A106, 1/F., No.109, Lijia Road, Henggang, Henggang  
Street, Longgang District, Shenzhen, Guangdong, China  
Tel: 0755-28680489  
E-mail: admin@ctnt-cert.com  
Web: www.ctnt-cert.com

Applicant's name ..... : Shenzhen Roundwhale Technology Co., Ltd.

Address..... : 202, 2/F., Building 27, Dafa Industrial Park, Longxi Community,  
Longgang District, Shenzhen, China

Manufacturer's name ..... : Shenzhen Roundwhale Technology Co., Ltd.

Address..... : 202, 2/F., Building 27, Dafa Industrial Park, Longxi Community,  
Longgang District, Shenzhen, China**Test specification:**Standard ..... : EN 60335-1:2012 + A11:2014 + A13:2017 + A1:2019 + A14:2019 +  
A2:2019 + A15:2021 + A16:2023

Test procedure ..... : Entrustment Test

Non-standard test method ..... : N/A

Test Report Form No. .... : IEC60335\_1Y

Test Report Form(s) Originator .... : Nemko AS

Master TRF ..... : Dated 2019-10-07

**Copyright © 2020 IEC System of Conformity Assessment Schemes for Electrotechnical Equipment  
and Components (IECEE System). All rights reserved.**

This publication may be reproduced in whole or in part for non-commercial purposes as long as the IECEE is acknowledged as  
copyright owner and source of the material. IECEE takes no responsibility for and will not assume liability for damages resulting from  
the reader's interpretation of the reproduced material due to its placement and context.

If this Test Report Form is used by non-IECEE members, the IECEE/IEC logo and the reference to the CB  
Scheme procedure shall be removed.

**This report is not valid as a CB Test Report unless signed by an approved IECEE Testing Laboratory  
and appended to a CB Test Certificate issued by an NCB in accordance with IECEE 02.**

**General disclaimer:**

The test results presented in this report relate only to the object tested.  
This report shall not be reproduced, except in full, without the written approval of the Issuing CTNT Testing  
Laboratory. The authenticity of this Test Report and its contents can be verified by contacting the CTNT,  
responsible for this Test Report.

**Remark:**



--

This report may not be reproduced in part without permission to avoid ambiguous interpretation.

This report can be checked and verified in the following ways.

Tel: 0755-28680489 E-mail: admin@ctnt-cert.com Web: www.ctnt-cert.com



<b>Test item description</b> ..... :	
<b>Product Name</b> ..... :	Combo Electrotherapy Device
<b>Trade Mark(s)</b> .....	N/A
<b>Model/Type reference</b> .....	M100A-WH1, R-W201A, R-W201B, R-W201C, R-W201D, R-W201E, R-W201, R-W101A, R-W101B, R-W101C, R-W101D, R-W101E, R-W101, R-C4D-RW1
<b>Ratings</b> ..... :	Charging input: 5V  0.3A Battery parameters: 3.7V  200mAh
<b>Test item particulars:</b>	
<b>Classification of installation and use</b> ..... :	Portable appliances
<b>Supply Connection</b> .....	Battery-powered
<b>Protection against electric shock</b> ..... :	Class III appliances
<b>Possible test case verdicts:</b>	
- test case does not apply to the test object .....	N/A
- test object does meet the requirement .....	P (Pass)
- test object does not meet the requirement .....	F (Fail)
- Not making judgments .....	-
<b>Testing</b> .....	
<b>Date of receipt of test item</b> ..... :	Aug.22, 2024
<b>Date (s) of performance of tests</b> .....	Aug.22- Aug.28, 2024
<b>General remarks:</b>	
"(See Enclosure #)" refers to additional information appended to the report. "(See appended table)" refers to a table appended to the report.	
Throughout this report a <input type="checkbox"/> comma / <input checked="" type="checkbox"/> point is used as the decimal separator.	
<b>Other important notes:</b>	
1. If you have any objection to the inspection results in this report, please submit a written report to the company within 15 days from the date of receipt of the report.	
2. Entrusting test only responsible for the incoming samples, and the test results are used by the entrusting party to understand the quality of the samples.	
3. This test report is invalid without the "test stamp".	
4. This report may not be reproduced in part without permission to avoid ambiguous interpretation.	
5. Test items with "*" are unauthorized items by CNAS.	
6. The remaining samples under test must be collected within three months of receipt of the inspection report. If the samples are not collected within the time limit, the laboratory will handle them by selves.	
Company name: Shenzhen Zhongwei Testing Technology Co., Ltd. Address: Room 1A106, 1/F., No.109, Lijia Road, Henggang, Henggang Street, Longgang District, Shenzhen, Guangdong, China Tel: 086-755-28680489 Email: admin@cntn-cert.com Website: www.cntn-cert.com	

This report may not be reproduced in part without permission to avoid ambiguous interpretation.



This report can be checked and verified in the following ways.

Tel: 0755-28680489 E-mail: admin@cntn-cert.com Web: www.cntn-cert.com



**Copy of marking plate:**

The artwork below may be only a draft. The use of certification marks on a product must be

Combo Electrotherapy Device  
Model: M100A-WH1  
Charging input: 5V  0.3A  
Battery parameters: 3.7V  200mAh  
Manufacture: Shenzhen Roundwhale Technology Co., Ltd.

Made in China

**General product information:**

- This appliance is a Class III appliances massager.
- This appliance is powered by batteries.
- The under test model M100A-WH1 and the series models R-W201A, R-W201B, R-W201C, R-W201D, R-W201E, R-W201, R-W101A, R-W101B, R-W101C, R-W101D, R-W101E, R-W101, R-C4D-RW1 are identical except model name and appearance decoration. The differences do not affect product safety and performance.

**Summary of testing:**

Determination of the result includes consideration of measurement uncertainty from the test equipment and methods.

According to the test of EN 60335-1:2012 + A11:2014 + A13:2017 + A1:2019 + A14:2019 + A2:2019 + A15:2021 + A16:2023.

This product is complied with the applicable requirements of standards.

**Testing procedure and testing location**

Laboratory name : Shenzhen Zhongwei Testing Technology Co., Ltd.

Testing location/address: : Room 106, 1/F, No. 109, Lijia Road, Henggang Community, Henggang Street, Longgang District, Shenzhen, Guangdong, China

Testing procedure : TL ☒ RMT ☐ SMT ☐ WMT ☐ TMP ☐

Tested By

(Test Engineer)

Mark Liu

Mark Liu

Reviewed By

(Supervisor)

Jackie Chen

Jackie Chen

Approved By

(Chief Engineer)

Flight Lee



This report may not be reproduced in part without permission to avoid ambiguous interpretation.

This report can be checked and verified in the following ways.

Tel: 0755-28680489 E-mail: admin@ctnt-cert.com Web: www.ctnt-cert.com