

TEST REPORT

REPORT NO.: CTNT240819014



Product name: Q-Switch Nd: YAG Laser System

Model No.: RD-YAG950B

Applicant: Shanghai Omni Laser Skinology CO., Ltd.

Test procedure: Type Test

Shenzhen Zhongwei Testing Technology Co., Ltd.



TEST REPORT**IEC 60825-1****Safety of laser products – Part 1:
Equipment classification and requirements****Report Number** : CTNT240819014**Date of issue** : December 12, 2024

Name of Testing Laboratory
preparing the Report..... : Shenzhen Zhongwei Testing Technology Co., Ltd.
 Room 1A106, 1/F., No.109, Lijia Road, Henggang, Henggang Street,
 Longgang District, Shenzhen, Guangdong, China
 Tel: 086-13662574496
 E-mail: admin@ctnt-cert.com
 Web: www.ctnt-cert.com

Applicant's name : Shanghai Omni Laser Skinology CO., Ltd.

Address : Building 12-A-6F, Liando U Valley Songjiang International Enterprise
 Port Park, No. 1698 Minyi Road, Songjiang District, Shanghai, China,
 201612

Manufacturer's name : Shanghai Omni Laser Skinology CO., Ltd.

Address : Building 12-A-6F, Liando U Valley Songjiang International Enterprise
 Port Park, No. 1698 Minyi Road, Songjiang District, Shanghai, China,
 201612

Test specification:**Standard** : IEC 60825-1: 2014**Test procedure** : Type Test**Non-standard test method** : N/A**Test Report Form No.** : IEC60825_1G**Test Report Form(s) Originator** : OVE**Master TRF** : Dated 2021-10-05**General disclaimer:**

The test results presented in this report relate only to the object tested.

This report shall not be reproduced, except in full, without the written approval of the Issuing CTNT Testing Laboratory.

The authenticity of this Test Report and its contents can be verified by contacting the CTNT, responsible for this Test Report.

Test item description : Q-Switch Nd: YAG Laser System**Trade Mark** : N/A**Model/Type reference** : RD-YAG950B**Ratings** : 100-240V~ 50/60Hz 10A

This report may not be reproduced in part without permission to avoid ambiguous interpretation.

This report can be checked and verified in the following ways.

Tel: 0755-28680489

E-mail: admin@ctnt-cert.com

Web: www.ctnt-cert.com

Responsible Testing Laboratory (as applicable), testing procedure and testing location(s):		
Laboratory Name	Shenzhen Zhongwei Testing Technology Co., Ltd.	
Testing location/ address.....:	Room 1A106, 1/F., No.109, Lijia Road, Henggang, Henggang Street, Longgang District, Shenzhen, Guangdong, China	
Tested by(Test Engineer)	Andrea Xie	
Reviewed by(Supervisor)	Oliver Long	
Approved by(Chief Engineer)	Flight Lee	
List of Attachments (including a total number of pages in each attachment): - Photo document - Test data graph		
Summary of testing:		
Tests performed (name of test and test clause): Determination of the result includes consideration of measurement uncertainty from the test equipment and methods. According to the test of IEC 60825-1: 2014 Safety of laser products – Part 1: Equipment classification and requirements. This product complied with the requirements of Class 4.	Testing location: Shenzhen Zhongwei Testing Technology Co., Ltd. Room 1A106, 1/F., No.109, Lijia Road, Henggang, Henggang Street, Longgang District, Shenzhen, Guangdong, China Tel: 086-755-28680489 E-mail: Web: www.ctnt-cert.com	
Test item particulars		
Classification of installation and use:	transportable / portable / stationary / mobile / fixed / permanently installed / hand-held	
Supply Connection.....:	internally powered / permanently installed / appliance coupler / non-detachable cord / detachable cord	
laser Class as defined in 3.44 and 3.18 through 3.24 of IEC 60825-1:2014.....:	Class 1, 1C, 1M, 2, 2M, 3R, 3B, or 4	
laser output (power or energy)	Maximum 1J (Adjusted 0.1-1J/cm ²)	
Accessories and detachable parts included in the evaluation	Footswitch	
Options included	N/A	

This report may not be reproduced in part without permission to avoid ambiguous interpretation.

This report can be checked and verified in the following ways.

Tel: 0755-28680489

E-mail: admin@ctnt-cert.com

Web: www.ctnt-cert.com